



**IMPACT AND USE OF POINT OF CARE HIV TESTING:
A PUBLIC HEALTH EVIDENCE PAPER**

Prepared by:

Mark Gilbert, MD

Division of STI/HIV Prevention and Control, BCCDC

November 29, 2007



BC Centre for Disease Control
AN AGENCY OF THE PROVINCIAL HEALTH SERVICES AUTHORITY

TABLE OF CONTENTS

EXECUTIVE SUMMARY	3
INTRODUCTION	5
METHODS	5
RESULTS	6
1. Characteristics of POC HIV tests compared to standard HIV testing.....	6
1.1 Test Performance.....	6
1.2 Window Period	6
1.3 Use of test kits.....	6
1.4 Provision of results	7
1.5 Quality control and assurance.....	7
2. Impact of Point of Care HIV testing compared to standard HIV testing	8
2.1 Test volume.....	8
2.2 Uptake of testing	8
2.3 Receipt of test results.....	8
2.4 Psychological impact.....	9
2.5 Time to availability of results and impact on medical decision-making	9
2.6 Entry into care	9
2.7 Impact on other testing	9
2.8 Impact on incident sexually transmitted infections	10
3. Settings and Indications for use of POC HIV tests.....	10
3.1 Testing in Pregnancy.....	10
3.2 Blood and body fluid exposures	11
3.3 Hospital-based HIV testing (Emergency rooms, inpatient admissions).....	11
3.4 STI or HIV testing clinics	12
3.5 Corrections settings.....	12
3.6 Outreach settings	13
DISCUSSION.....	13
TABLE 1: Description of studies assessing the impact of POC HIV Testing	15
REFERENCES	19

EXECUTIVE SUMMARY

Point of care (POC) HIV tests are screening tests for HIV available for use by health care professionals in a clinical setting, and are licensed for use in Canada. This paper reviews the evidence regarding use and impact of POC HIV tests, in order to determine the best use of this technology in BC.

The following key characteristics of POC HIV tests were identified:

- POC HIV tests licensed for use in Canada will have similar sensitivity and specificity compared to standard HIV screening tests ($S_n, S_p > 99\%$). False positive POC HIV tests will occur (particularly in settings with low HIV prevalence)
- With standard HIV testing, the result returned to the patient is final. While a negative POC HIV test is considered final and no confirmatory testing is required, a positive POC HIV test is a preliminary result and confirmatory testing is required
- The window period may differ slightly between POC HIV test products and standard HIV screening tests, which may lead to infrequent discrepant results
- Typically health care professionals find POC HIV tests to be easy to use. Unlike automated protocols in place for interpretation of standard HIV testing, interpretation of POC HIV tests is subjective. Inter-reader variability in test interpretation is low, although variability may be greater in early HIV infection
- With standard HIV testing, a follow-up visit is required for receipt of results. The same applies to POC HIV testing if the result is preliminary positive; however if the test is negative a follow-up visit is not required
- Unlike standard HIV testing, the health care professional administering the POC HIV test assumes the responsibility for quality assurance activities to ensure that the test is carried out correctly

The potential impacts of POC HIV testing are as follows:

- Use of POC HIV testing may result in increased acceptance and volume of HIV testing
- Individuals undergoing POC HIV testing are more likely to receive their test result, particularly if HIV negative. Receipt of a final HIV positive result may not differ from standard testing, although individuals may be more likely to present independently for receipt of confirmatory test results
- While not well evaluated, it is possible that there may be a negative psychological impact for some individuals who receive a preliminary positive or false positive POC HIV test result
- The rapid turnaround time associated with POC HIV testing can guide urgent decision-making to prevent HIV infection or to improve patient care
- POC HIV testing may lead to decreased uptake of testing for other infections (e.g., hepatitis C, syphilis)
- Increased incidence of sexually transmitted infections has been associated with the use of POC HIV testing, possibly due to disinhibition on receipt of a negative test result or compression of counselling into a single visit

POC HIV tests have been recommended for testing of pregnant women with unknown or undocumented HIV status at the time of labour, and for testing of source or exposed individuals in blood and body fluid exposures (particularly for occupational exposures). The use of POC HIV tests has also been found to be feasible in the following settings: hospital-based settings (e.g., emergency rooms, inpatient wards), STI or HIV testing clinics, corrections facilities, and a variety of different outreach programs.

The findings of this review demonstrate that POC HIV tests may be of use in BC, however, their potential benefit must be weighed against potential harms. Studies reviewed here were largely based in large, urban centres in the US in settings with high volumes of HIV testing, which may limit the generalizability of these findings to BC.

The development of new test technologies such as POC HIV testing may facilitate the identification of persons infected with HIV and their engagement in HIV prevention and connection to appropriate care and support. When considering the use of a POC HIV test product, these findings should be weighed against the specific characteristics of the POC test product and standard HIV testing protocols in place.

INTRODUCTION

Point of Care (POC) or rapid HIV antibody tests are screening tests for HIV which are available for use in a clinical setting, providing results typically within minutes. The first POC HIV tests were developed in the late 1990's, and kits have been developed for use with oral fluid, fingerstick or venous blood specimens. In the US these products have been adopted for use in multiple settings, with the development of a national rapid test distribution program in 2003.¹ In Canada, these kits have had limited use as a previously available product was removed from the market in 2002 due to concerns regarding test quality.² With a new POC HIV test recently licensed³ and the potential for future test products to be marketed in Canada, it is timely to consider the impact of POC HIV testing in order to determine the best utilization of this technology.

While guidelines regarding the use of POC HIV test kits and accompanying counseling have been developed^{4,5}, the potential public health impact of these tests is uncertain. Of particular interest is the ability of POC HIV tests to contribute to provincial and regional goals for the prevention of new HIV infections, reducing the number of HIV positive individuals who are unaware of their HIV status, and linkage of HIV positive individuals to care, treatment and support services.

The purpose of this evidence paper is to provide the context required to guide decision-making by health officials regarding the adoption of POC HIV tests in BC. This review set out to answer three questions:

- i) What are the characteristics of POC HIV tests compared to standard HIV testing?
- ii) What is the impact of POC HIV tests compared to standard HIV testing?
- iii) In what settings and for what indications has POC HIV testing been used or recommended?

Please note that this is a general review of POC HIV tests. Specific characteristics or impacts may differ depending on the POC HIV test and specimen type under consideration and the current standard HIV testing protocols in place. These need to be considered in evaluating the potential regional impact of specific POC HIV test products. As new standard HIV test protocols are implemented (such as pooled nucleic acid amplification testing (NAAT) or fourth generation enzyme-linked immunoassays (ELISA)) the impact of an adopted POC HIV test product may need to be reconsidered.

METHODS

For this review, we considered POC HIV tests to be HIV tests which provide a result within minutes to hours, with same day provision of test results at the point of care (i.e., at the site of interaction with client being tested such as within a clinic or in an on-site laboratory). We defined standard HIV testing as HIV antibody testing requiring venipuncture and follow-up for receipt of results (i.e., a screening test such as a second or third generation ELISA and a confirmatory test as required such as a Western Blot).

We searched MEDLINE using the subject heading Point-of-Care systems, and the key words HIV, test, testing, rapid, point of care. Articles describing the use or impact of POC HIV tests prior to August 2007 were identified and reviewed, and relevant citations from bibliographies were retrieved. Online abstract databases for scientific conferences related to HIV were also

searched for relevant material (i.e., International AIDS conferences, Conferences on Retroviruses and Opportunistic Infections, Canadian Association for HIV Research).

We included studies from industrialized countries only as these studies are more likely to be generalizable to British Columbia. Tests for all specimen types were included. Studies which were designed primarily to evaluate the diagnostic properties of specific POC HIV test products were not included. For the assessment of the impact of POC HIV testing compared to standard HIV testing, we reviewed studies which included a comparison between POC HIV testing and standard HIV testing, excluding results which could be biased by differences between comparison arms (e.g., using different recruitment approaches).

RESULTS

1. Characteristics of POC HIV tests compared to standard HIV testing

1.1 Test Performance

Any POC HIV test licensed for use in Canada will have similar performance compared to standard HIV screening tests (sensitivity and specificity > 99%).⁶ POC HIV tests have high negative predictive value and a negative result is considered final; however, the positive predictive value will vary according to HIV prevalence and false positive results will occur. Consequently a positive POC HIV test result is considered a preliminary result which requires confirmation by collection of a venous blood specimen. Some have recommended qualifying a preliminary positive result according to an individual's risk of HIV (i.e., as a positive POC test result in a person from a population with a high prevalence of HIV is more likely to be a true positive).⁷ With standard HIV testing, screening and confirmatory testing if required are performed in sequence, and the result returned to the client is considered final.

In a summary of the US rapid test distribution program in 2003-2005 of over 370,000 OraQuick POC HIV diagnostic tests, 5385 (1.4%) of tests were preliminary positive, of which 4640 (1.2%) were confirmed positive.¹

1.2 Window Period

The window period may differ slightly between POC HIV test products and the standard HIV screening tests in use which may lead to discrepant results, and may indicate differing sensitivities for detecting early infection.⁸⁻¹⁰ In one evaluation of POC HIV test products on established HIV seroconversion panels, POC tests demonstrated seroconversion at equal or slightly later times to standard HIV screening tests.⁸

Like standard HIV testing, POC HIV test products are unable to detect acute HIV infection prior to seroconversion.¹¹ In San Francisco and Seattle, nucleic acid amplification testing (NAAT) is performed on blood specimens from high risk individuals with negative POC HIV tests in order to identify individuals with acute HIV infection (approximately 1% of individuals undergoing POC HIV testing, increasing case detection by up to 18%).¹²⁻¹⁴

1.3 Use of test kits

POC HIV tests are generally found by health care providers to be easy to use.⁹ Use typically involves collection of an appropriate specimen then following test kit instructions (e.g., addition

of reagents, a series of timed steps) to achieve a result. Test kits usually include an internal control which assists in appropriate interpretation of results by identifying tests which have been performed incorrectly.

POC HIV tests are typically used by health care providers who do not have formal laboratory training and require subjective interpretation of visual results (in contrast to standard HIV testing which is performed in a laboratory and uses automated protocols for interpreting screening tests). Few studies have looked at inter-reader variability in test interpretation. A World Health Organization assessment of a variety of POC HIV tests by three independent readers on defined HIV positive and negative panels demonstrated low inter-reader variability (range 0 to 4.6% per product).⁸ A US study examined the interpretation of a POC HIV test kit by untrained health care workers from a variety of disciplines (following manufacturer's instructions only) on patient specimens and evaluation panels using sera with differing strength of antibody signals.¹⁵ Excluding invalid test results (<10% of tests performed, due to incorrect procedure and discounted), of 856 tests 96.6% were interpreted correctly (range 93-100% per discipline). Of the remainder, 1.9% were falsely read as negative and 1.5% were falsely read as positive. Variability in POC test result interpretation may be greater in early HIV infection.¹²

1.4 Provision of results

One of the attractive features of POC HIV tests is the same-visit availability of a test result, and if the test is negative and exposure did not occur within the window period no additional testing is necessary. This differs from standard testing where there is a delay between specimen collection and result availability (in some jurisdictions up to two weeks after specimen collection). Additionally pre- and post-test counseling are incorporated within the same, initial visit and must be tailored to the use of POC HIV tests.^{4,16}

1.5 Quality control and assurance

Unlike standard laboratory-based HIV testing, the health care professional administering the POC HIV test assumes the responsibility for ensuring that the test is carried out correctly. Sites adopting POC HIV testing are recommended to implement appropriate quality assurance measures, including staff training, documentation and monitoring of test outcome, and regular use of test control kits.¹⁶⁻¹⁹ Through these measures, regional POC HIV testing programs have been able to monitor POC HIV test performance and identify and investigate potential problems. For example, in 2004-05 in the US several regions identified increases in false positive results using an oral fluid POC HIV test at specific test sites.²⁰⁻²³ While subsequent investigations failed to identify any explanation for these results and the lower specificity at certain sites was determined to be within the acceptable range^{21,22}, site-specific factors related to test use and interpretation could not be ruled out leading some jurisdictions to switch to a different testing protocol.²⁴

In addition to ensuring POC test kits are performing optimally at the site of use, it is important to ensure consistent quality of POC test kits and that their performance is within licensed specifications. In BC an unexpected increase in false negative results using a POC HIV test product in 2002 and reduced test sensitivity was attributed to changes in the manufacturing process, leading to the withdrawal of the product from market and recommendations for retesting of individuals with negative test results.² It is recommended that a robust quality assurance process be in place prior to widespread adoption of a specific POC HIV test product.²⁵

2. Impact of Point of Care HIV testing compared to standard HIV testing

We identified 25 studies that compared POC rapid HIV testing to standard HIV testing (Table 1).^{9,26-49} The majority of studies were observational in design (20 studies), including 17 studies evaluating POC HIV testing programs with comparison to a baseline period where standard HIV testing was offered, and three studies using concurrent controls (patients choosing standard HIV testing over rapid HIV testing, or simultaneous standard HIV testing). The remaining non-observational studies included one meta-analysis and four randomized controlled trials.

2.1 Test volume

Five studies reported an increase in test volume following introduction of POC HIV testing.^{26,28,31,39,46} In only one study was the use of POC HIV testing not accompanied by promotion efforts or a different recruitment strategy from that used during the baseline period.²⁶ In this large study of 61 HIV testing clinics total HIV test volume increased by 36.9% following introduction of POC HIV testing (with >99% of clients choosing rapid testing).

2.2 Uptake of testing

There is good evidence to suggest that rapid HIV testing can result in increased HIV test uptake or acceptance in outreach settings. In one RCT among MSM in bathhouses and needle exchange clients, rapid HIV testing was significantly associated with 1.3 to 1.7 times increased acceptance of testing respectively per site compared to standard HIV testing.⁴⁷ Smaller studies in other settings also report increased test uptake.^{44,48}

The impact of POC HIV testing on engaging individuals who have not previously tested was not assessed by any study reviewed. Where reported the proportion of participants without a history of previous testing ranged from 9-28% in outreach or clinic settings^{42,47,49} up to 55% in one emergency room.³⁶ In one study, 11% of clients undergoing rapid HIV testing at anonymous test sites reported that they would not have tested if rapid HIV testing were not available.²⁶

2.3 Receipt of test results

It is commonly reported that with standard HIV testing a proportion of both HIV negative and positive clients will not return for test results; this proportion is variable by test site and population testing (ranging from 16 to 58% in some studies).⁵⁰⁻⁵⁴ POC HIV testing allows for same-visit delivery of test results, with follow-up for confirmatory test results only required for those who obtain a preliminary positive result. Improved receipt of test results is commonly cited as an important reason for adoption of POC HIV tests. Receipt of test results is important as knowledge of HIV positive status is associated with a reduction in self-reported risky sexual behaviours⁵⁵ (the impact of receipt of a negative HIV test is less understood). Site-specific HIV prevalence and return rate for response should be considered in assessing this impact of POC HIV testing.⁵⁶

There is strong evidence that use of POC HIV testing improves the receipt of HIV test results. Studies in multiple settings and of varying design have consistently demonstrated that individuals tested by POC HIV tests have a greater receipt of final test results (ranging from 95-100% of participants overall).^{9,26,28,37,39,41-43,47-49} In a recent meta-analysis of the efficacy of

alternative HIV testing and counseling approaches in improving the receipt of results, the use of rapid HIV testing significantly increased receipt of test results (RR 1.80). The greatest impact was observed in emergency room settings, followed by STI clinics and outreach venues, and the overall increase in receipt of rapid test results was greater for HIV negative (RR 2.00) than HIV positive persons (RR 1.19).³³

In this meta-analysis receipt of a preliminary positive result in HIV positive persons was the outcome of interest, and some have thought that the receipt of a preliminary positive POC HIV test result and accompanying counseling alone may be of benefit. A relevant question is the impact of POC HIV testing on receipt of final confirmed results in HIV positive individuals (including individuals with false preliminary positive HIV results). The evidence is scant on this issue, likely due to the overall small number of HIV infections identified per study. Some studies have reported an increased receipt of final test results among HIV positive persons with rapid testing^{28,35,41} while other studies find no significant difference.^{26,36} Individuals who receive a preliminary positive result by rapid testing may be more likely to return of their own volition for receipt of the confirmatory test result, and less likely to require finding by public health staff.^{26,35} In the US national rapid test program in 2003-2005, 21% of individuals who received a preliminary positive POC HIV test result did not return for receipt of confirmatory test results.¹ It is important to note that individuals with false preliminary positive results may also not return for receipt of confirmatory test results.

2.4 Psychological impact

The psychological impact of POC HIV testing compared to standard HIV testing for patients undergoing testing has not been assessed. Two studies evaluating the testing of source patients during occupational blood exposures identified decreased anxiety among exposed staff when POC HIV testing was used.^{38,40}

2.5 Time to availability of results and impact on medical decision-making

The use of POC HIV tests results in a significantly reduced turnaround time for results (dependent on the POC HIV test kit used).^{27,29,39,40} Where knowledge of HIV status is required to guide urgent medical decision-making (for clinical care or prevention of HIV infection) POC HIV testing may be superior. The introduction of POC HIV testing for source patients in the evaluation of occupational blood exposures for health care workers has been found to result in fewer initiations or shorter courses of antiretroviral post-exposure prophylaxis.^{38,40,44} Similarly, the use of POC HIV testing in labour and delivery units for women with undocumented HIV status has been associated with increased availability of results prior to delivery and greater opportunity for administration of antiretroviral prophylaxis to HIV-positive women.^{29,32}

2.6 Entry into care

Few studies have assessed the impact of POC HIV testing on entry into HIV clinical care for persons with newly identified HIV infection. Two studies assessing this outcome obtained conflicting results. While one study found that the introduction of POC HIV testing led to an increased proportion of HIV-positive individuals presenting for HIV clinical care, no significant difference was found in the second study.^{28,36}

2.7 Impact on other testing

POC HIV testing may have a negative impact on testing for other infections (e.g., Hepatitis C, STI) although this has not been well assessed. In Houston, a decline in the number of syphilis tests performed by community-based organizations was observed following the implementation of a POC HIV testing program (from 710 to 41 tests per month, a 95.7% decrease), leading to a recommendation to re-emphasize the use of venipuncture as the primary source of testing for HIV.³⁴ Similarly, following the introduction of POC HIV testing in a public health clinic in Japan a decrease in the proportion of clients tested for syphilis (from 77% to 63%) and chlamydia (from 76% to 33%) was observed; however, the overall number of tests for syphilis and chlamydia increased due to the increased client volume attributed to offering POC HIV testing.⁴⁵

2.8 Impact on incident sexually transmitted infections

An early observational study of POC HIV testing in STI and anonymous HIV testing clinics did not demonstrate an increase in STI incidence following introduction of POC HIV testing (based on return of clients with a new STI).³⁵ A recent RCT in an STI clinic setting was designed to assess the relative efficacy of POC HIV testing to standard testing on STI incidence.⁴² In this trial, clients randomized to POC or standard HIV testing were followed over 12 months with repeated specimen collection for STI (including HIV). At each follow-up time point the overall incidence of STI was higher among patients who received POC HIV testing (19.1% vs 17.1% at 12 months); this difference was significant for males (15.5% vs 11.6%, RR 1.34 at 12 months) and individuals without STI detection at baseline (10% vs 7%, RR 1.44 at 6 months). There was also a trend towards an increased STI incidence among MSM clients (21.8% vs 11.8% in heterosexual men).

This study concluded that in the short term and in some sub-groups, rapid HIV test interventions may be less effective at preventing STI, and that the use of POC HIV tests may be most indicated in settings with high prevalence of HIV infection and a low rate of return for HIV test results. Potential explanations for these findings included disinhibition following receipt of a negative test result, and slightly greater time with a counselor (development of risk reduction plan, reinforcement of prevention messages) in the standard testing group over two visits.

3. **Settings and Indications for use of POC HIV tests**

3.1 Testing in Pregnancy

US and UK guidelines recommend the use of POC HIV testing for pregnant women with unknown or undocumented HIV status at the time of labour.^{57,58} Nine studies have evaluated the use of POC HIV tests in pregnant women with undocumented HIV status (where eligibility for testing is usually defined as gestational age \geq 34 weeks, or active labour at gestational age $>$ 24 weeks; ranging from 8-50% of women seen).^{27,29,30,32,59-63} Acceptance of POC HIV testing among eligible women was high (range 69-95%)^{27,30,32,59-61}. All studies successfully identified women with HIV, with the majority of women (range 63-76%)^{30,32,61,62} and exposed infants (range 89-100%)^{30,61,62} receiving ART prior to delivery or after birth, respectively. However, in most studies women with false positive POC HIV test results were also identified (range 0.07-1.5% of women)^{27,29,30,32,60,63}, leading in some cases to unnecessary ART.^{27,63} Clinical judgment, knowledge of HIV prevalence and POC HIV test characteristics are required to counsel pregnant women appropriately.^{27,60}

The main concern expressed about POC HIV testing in pregnant women at the time of labour is related to the capacity for informed consent.⁶⁴ No studies have assessed the experience or

attitudes of women with undocumented HIV status at the time of delivery to POC HIV testing. Three pilot studies have looked at this question with women at low risk for HIV or with a high compliance with prenatal HIV testing; in these studies, the majority of women felt that POC HIV testing was important and should be offered during pregnancy.⁶⁵⁻⁶⁷ In one pilot study, all pregnant women presenting in labour at an inner-city hospital were offered POC HIV testing.⁶⁵ The majority of women in this study demonstrated good understanding of test rationale and felt reassured by the test, with 80% stating it was not difficult to make a decision about HIV testing (14% reported difficulty due to contractions/pain). A minority of women reported feeling insulted or angry (9%), high anxiety (4%) or embarrassed (4%) at POC HIV test offer.

In a study piloting an informed consent form for POC HIV testing for women in labour (39% in active labour, 61% reporting substantial pain), the majority were able to restate the purpose and benefits of testing (to prevent vertical transmission), although understanding of the potential for unnecessary ART treatment if false positive was initially poor.⁶⁶ Similar findings were found in another pilot of an informed consent form in Canadian women who had given birth in the previous five years.⁶⁷

POC HIV testing has also been incorporated into prenatal visits. In a UK study, the majority of low-risk women with a high compliance with HIV testing in prenatal care indicated that POC HIV testing would be acceptable, including approximately a third of women who did not accept standard testing during their pregnancy.⁶⁸ An evaluation of POC HIV testing in women in prenatal care in the US verified this finding, with high test acceptance (72%).⁶⁹

3.2 Blood and body fluid exposures

HIV testing of source and exposed individuals is recommended in evaluation of blood and body fluid exposures to HIV.⁷⁰⁻⁷⁴ In occupational settings, the use of POC HIV testing and timely receipt of test results is acknowledged to facilitate decision-making regarding use of post-exposure prophylaxis (PEP).^{57,70-72}

No studies have evaluated the use of POC HIV testing in sexual exposures, and in this scenario the source may be less available for testing. However, HIV testing of exposed persons is recommended to avoid unnecessary PEP in individuals already infected with HIV.^{70,72} In one case report, unnecessary PEP was avoided during an evaluation of a sexual exposure when the exposed individual was determined to be HIV positive.⁷⁵

3.3 Hospital-based HIV testing (Emergency rooms, inpatient admissions)

There has been considerable attention to the issue of HIV testing in hospital settings given the recognition that marginalized HIV positive individuals unaware of their HIV status may access hospital care prior to diagnosis⁷⁶ and recent recommendations for opt-out screening in this setting.⁵⁸ A number of studies in large urban centres have looked at the use of POC HIV testing in emergency room, inpatient wards or outpatient clinic settings – usually in the context of evaluating universal screening compared to targeted testing.^{9,28,36,39,48,77-84} Consequently it is difficult to separate the impact of POC HIV testing from the impact of differing recruitment strategies which are not the focus of this paper.

In general, the use of POC HIV tests is feasible in these settings, with programs demonstrating an acceptance of testing ranging from 24-98% among individuals not known to be HIV positive or to have recently tested. The most commonly reported reasons for refusal of testing include

having tested recently or not perceiving self at risk for HIV.^{36,48,79,81,83} There is limited information regarding the experiences of individuals undergoing rapid HIV testing in hospital settings. A qualitative assessment of emergency room patients offered rapid or standard HIV testing identified a general preference for rapid testing due primarily to the increased turnaround time for results.⁸⁵ The infrastructure for POC HIV testing in hospital settings must be in place, as provider barriers to conducting testing can include ignorance regarding test protocols, and time required to obtain informed consent and perform testing.⁸⁴

While not evaluated in any of the identified studies, POC HIV testing may be of use in hospital settings in patients where knowledge of HIV status will have an impact on further workup or treatment (e.g., in patients presenting with opportunistic infections). A common problem with standard HIV testing in the hospital setting is the delay to result availability, which may affect therapeutic choices. In addition, using standard HIV testing results may not be available until after discharge and public health case follow-up is difficult. The use of POC HIV testing for diagnosis of HIV in hospital settings may be superior to standard testing as individuals are more likely to receive their results at the same visit, and there may be greater opportunities to facilitate public health follow-up and early connection to clinical care. One retrospective chart review comparing the outcomes of inpatients tested by POC or standard HIV testing identified reduced time to diagnosis and admission to the inpatient HIV ward, and to first outpatient visit for HIV-related care in patients undergoing POC HIV testing.⁸⁶

3.4 STI or HIV testing clinics

When given a choice, 90% or more of clients attending STI or HIV testing clinics for HIV testing will choose rapid HIV testing and same-visit result availability over standard testing.^{26,49,87,88} While up to 30% of clients report that POC HIV testing is stressful, the majority of clients prefer POC HIV testing and same-day delivery of results.^{26,35,42,49,87-89} Clients choosing or preferring standard testing commonly report valuing a confirmed or more accurate result, perceiving less stress related to testing, and more time to prepare for a result (particularly among first-time testers).^{26,42,88,89} Clients presenting to STI or HIV testing clinics should be given a choice of preferred testing options.

Typically clinic or program staff report initial apprehension about using POC HIV tests; however, after appropriate training and after a short period of use the majority of staff report feeling very comfortable with the use of these tests.^{26,35,88,90} Concerns include lack of time to prepare for and discomfort with delivering a positive result, client perceptions of accuracy of result, and reduced time for counseling and reinforcement of prevention messages without a second visit. Perceived advantages include continuity of counseling, ensuring receipt of test results, greater ability to focus on risk issues, and convenience.

3.5 Corrections settings

A common priority area for HIV testing are corrections facilities, which may be an ideal point of engagement in HIV testing for individuals with undiagnosed HIV infection. It has been estimated that a substantial portion (up to 26% in 1997) of the population with prevalent HIV infection passes through the corrections system each year in the US.⁹¹ The 1994 prevalence of HIV in BC provincial prisons was estimated at 1.1%.⁹²

Two studies have reported evaluations of POC HIV testing programs in US Corrections facilities.^{80,93} The largest is a recent report of a four state rapid HIV testing program in the US.⁹³

Programs were set up at each facility in collaboration between corrections programs and state health departments, with promotion of HIV testing to male and female inmates; inmates were tested if they requested HIV testing or were referred for testing by medical staff. Over a 2.5 year period, 6% of 550,000 new inmates were tested of which 1.3% of individuals had a preliminary positive result (1.2% confirmed positive). The other reported approach is the routine offer of HIV testing to new admissions, which in a female corrections facility resulted in 46% of 2128 inmates accepting HIV testing.⁸⁰ In both studies, preliminary POC HIV test results were received by >99% of inmates tested. A third pilot feasibility study, in a corrections facility with a strong history of inmate HIV testing, also found POC HIV testing to have a high acceptance rate.⁹⁴

In these studies, many individuals with new HIV infections did not report risk factors for HIV infection^{80,93} and in one study individuals who reported no HIV risk factors were four times more likely to be HIV positive.⁹³ Disclosure of risk behaviour in corrections may be challenging, and risk-based testing may be less successful.⁹⁵ Other challenges include delivery of confirmatory test results to released individuals^{93,94}, connection to HIV clinical care⁸⁰, and difficulty in determining the best time for testing (i.e., balancing potential distress or intoxication at admission with rapid turnover).^{94,95}

3.6 Outreach settings

Incorporating POC HIV testing into outreach programs is attractive due to both the ability to engage untested or high risk persons in HIV testing and same-day provision of test results in populations which may not return for a follow-up visit. Numerous studies have demonstrated the feasibility of POC HIV testing programs in outreach settings targeting high risk populations, including: bathhouses⁴⁷, gay pride events⁹⁶, needle exchange sites^{47,97}, sex work venues and programs^{98,99}, homeless shelters and single room occupancy hotels^{97,100}, community centres accessed by youth¹⁰¹, ethnic minority groups^{102,103}, mobile clinics^{41,104}, and substance abuse programs/detoxification centres (community or hospital-based).^{97,100,105,106} These programs have demonstrated variable acceptance of testing by site (ranging from 14-82% of persons approached). Concerns have been identified with some of these approaches (e.g., confidentiality, privacy, clients not prepared for testing, testing leading to disinhibition or unprotected sex).^{64,107,108} These concerns have not been well evaluated.

While the use of POC HIV testing improves overall receipt of HIV test results, individuals in some outreach settings who receive a preliminary positive HIV result may not agree to or return for the results of confirmatory testing.¹⁰⁰ Some clients may also choose not to receive a preliminary result of a POC HIV test.⁹⁹

DISCUSSION

The findings of this review demonstrate that POC HIV tests have merit, and their use has been demonstrated to be feasible in a variety of settings. POC HIV testing is highly acceptable to and preferred by patients and staff performing testing in most settings, and introduction of POC HIV test programs may result in increased uptake and volume of HIV testing. Individuals undergoing POC HIV testing are more likely to receive their test result, particularly individuals who are HIV negative, which may be particularly useful in settings where receipt of results is difficult such as some STI clinics and outreach programs. While the benefit on final receipt of confirmed HIV positive results is less clear, studies have reported equivalent or increased receipt compared to standard testing, and individuals receiving a preliminary positive POC HIV test result may be

more likely to return for a confirmatory test result, reducing public health follow-up efforts. Improved receipt of test results and timely connection to public health follow-up and clinical care may also be of benefit in settings such as emergency rooms or corrections facilities. Finally, the rapid availability of results can facilitate more timely prevention of vertical transmission and avoidance of unnecessary PEP to prevent occupational transmission of HIV.

As with the potential introduction of any new health technology, these merits need to be weighed against disadvantages and potential harms. POC HIV testing by definition is not laboratory-based, and with shift of testing to the point of care and subjective interpretation of results there may be greater potential for user error or other site-specific factors to influence the quality of testing. Recommendations for quality assurance measures, staff training, and documentation and monitoring of test outcomes must be heeded in any site where POC HIV tests are used. The use of POC HIV tests may have a negative impact on the number of individuals tested for other infections (e.g., Hepatitis C, syphilis), and of even greater concern is the potential for increased risk of STI. Finally, with POC HIV testing, individuals will receive preliminary positive or false positive results and may require additional support. These potential harms have not been well evaluated. Other issues of concern such as impact of POC HIV testing on case reporting to public health and HIV surveillance, and monitoring of HIV testing volume and patterns have not been assessed.

There are limitations to this review that may affect the generalizability of these findings to BC. The majority of these studies are from the US, where important differences exist in the populations affected by HIV, in access to free HIV testing, and the regulatory framework surrounding HIV testing. All studies were from urban, typically inner-city settings, and usually in settings where large volumes of HIV testing were performed (no studies in rural or remote settings were identified). Finally, these studies spanned a decade of research, a variety of different POC HIV test products, and a number of different program models.

The development of new test technologies such as POC HIV testing provides new opportunities to identify persons infected with HIV, engage them in HIV prevention and connect them to community resources, clinical care and effective treatment. When considering the use of a POC HIV test product, the findings in this report must be weighed against the specific characteristics of the POC test and standard HIV testing protocols in place. The studies reviewed here have compared POC HIV testing to standard HIV testing protocols which include second or third-generation ELISA HIV screening assays (similar to what is currently in use in BC). The role and impact of POC HIV testing must also be weighed against other new HIV testing technologies which may become available in the near future (e.g., fourth-generation ELISA or NAAT) which have the capacity for increased identification of persons acutely infected with HIV.

TABLE 1: Description of studies assessing the impact of POC HIV Testing

Study (Period, Location)	Setting	Study design	Intervention (POC test used)	Control	Eligibility	Sample size (Intervention arm)	Outcome	Result (Intervention compared to Control arms)	Ref
Hutchinson 1990-2005	Various	Meta-analysis	Rapid HIV test	Standard HIV test	---	17520 (NS)	Result receipt	Increased return for results (overall RR 1.80 [1.46, 2.22]). Less pronounced in studies based on RCT or concurrent (vs baseline) controls. Greatest impact in ER (RR 2.19) over STD or outreach, and HIV negative (RR 2.0) over HIV positive (RR 1.19).	³³
Metcalfe 1999-2000 USA	STI clinic	RCT	Rapid HIV test (SUDS)	Standard HIV test	HIV-vaginal/anal sex in past 3 months 15-39 yrs	3342 (1648)	STI incidence Result receipt	Increased STI incidence in intervention arm (19.1% vs 17.1% at 12 mos, not sig). Significant increase in males (RR 1.34 at 12 mos) and participants without STI at baseline (RR 1.44 at 6 mos). Trend to increased rate in MSM (<10% of sample). Increased receipt of test result (99% vs 69.4%).	⁴²
Spielberg 1999-2000 Seattle	NEX, bath-house	RCT	Rapid HIV test (SUDS)	Standard HIV test. Option of getting results by phone.	Age>14 HIV-last test > 3 mos	3455 (1633)	Test acceptance Result receipt	Increased acceptance: NEX OR 1.7, bathhouse OR 1.3 Increased receipt of test results: NEX OR 2.4, bathhouse OR 1.9	⁴⁷
Wurcel 2003-2004 Boston	Hospital (in-patient)	RCT	Rapid HIV test (OraQuick)	Standard HIV test	In/outpatient Physician-referred HIV-last test > 1 mos	203 (101)	Test acceptance Result receipt	Trend towards higher acceptance of testing (59.4% vs 41.2%), significant in Caucasian, Hispanic subgroups. Greater proportion received results (p<0.001).	⁴⁸
King 1998-1999 Switzerland	Occupational exposure	RCT	Rapid HIV test of source (Genie II)	Standard HIV test of source	Source patients from occupational blood exposure	60 (30)	False Negative results PEP uptake Psychological distress	No false negative results identified Lower use of PEP (3% vs 83%; p<0.001) Increased proportion reporting no anxiety or stress (24% vs 7%)	³⁸
Kelen 1994-1995 Baltimore	ER	OBS (baseline control)	Rapid HIV test (SUDS)	Standard HIV test	18-55 years HIV-	1449 (467)	Result receipt Entry to HIV care	No significant difference among HIV+ (73% vs 62%). No significant difference in follow-up for HIV clinic appointment (55% vs 64%) (NB: includes 146 persons who were given a choice between rapid and standard testing)	³⁶
Kendrick 1999-2000 Chicago	STI clinic	OBS (concurrent control)	Rapid HIV test (SUDS)	Standard HIV test	HIV-Last test > 3 mos Not presenting for results	1581 (1372)	Result receipt Entry to HIV care	Increased receipt (>99% vs 84%) Greater entry into care (97% vs 84%) and retention after 1 month (83% vs 68%). (NB: possible differences in ensuring compliance)	³⁷

Study (Period, Location)	Setting	Study design	Intervention (POC test used)	Control	Eligibility	Sample size (Intervention arm)	Outcome	Result (Intervention compared to Control arms)	Ref
Liang 2003-2004 Baltimore	Mobile clinic	OBS (concurrent control)	Rapid HIV test (Oraquick)	Standard HIV test	Recruited on street or presenting for testing	439 (284)	Predictors of test choice Result receipt	Rapid test choice associated with African-American ethnicity (OR 1.91), drug treatment in last three mos (OR 0.58). Not CSW, IDU, STI history. Increased receipt of post-test counseling (HIV- 93% vs 40%, HIV+ 89% vs 11%)	⁴¹
Guenther 2001-2002 Toronto	HIV testing clinic	OBS (concurrent control)	Rapid HIV test (Fast Check)	Standard HIV test	NS	1610 (1468)	Predictors of test choice Receipt of test results	Rapid test choice significantly less likely in females (27% vs 39%), more likely in first testers at clinic (43% vs 32%), and persons aware of choice of test prior to arrival at clinic (34% vs 26%). No difference in risk profile. Increased receipt (99.7% vs 93.4%). Of HIV+, final result received by 100% and 82% of rapid and standard testers respectively.	⁴⁹
Jackson 2005 Houston	CBO	OBS (baseline control)	Rapid HIV test (Unknown)	Standard HIV test	NS	NS (NS)	STI testing	Decrease in syphilis testing by 95.7% after introduction of rapid testing (from 710 to 41 specimens per month)	³⁴
Kassler 1993 Dallas	STI clinic, HIV testing clinic	OBS (baseline control)	Rapid HIV test (SUDS)	Standard HIV test	NS	4555 (2477)	Result receipt STI incidence	Increased receipt (HIV- 96% vs 59%, HIV+ 98% vs 88%). HIV+ more likely to return for confirmatory result without field follow-up (94% vs 45% at STI clinic site). No significant difference in STI incidence at clinic over 1 year (6.0% vs 5.9%)	³⁵
San Antonio-Gaddy 2003 New York	HIV testing clinics	OBS (baseline control)	Rapid HIV test (Unknown)	Standard HIV test	NS	10595 (6122)	Test volume Result receipt	Increased volume: 36.9% increase compared to baseline (p<0.001; of which >99% chose rapid test) Increased receipt (HIV- 99% vs 84%, p<0.0001; HIV+ 75% vs 72%, not sig). Greater return for receipt of confirmatory test result (81% vs 72%).	²⁶
Shima 2003-2004 Japan	HIV testing clinics	OBS (baseline control)	Rapid HIV test (Determine HIV-1/2)	Standard HIV test	NS	NS	Test volume	Increased test volume at each clinic site (350%, 1200%), with promotion of testing.	⁴⁶
Henn 2005 New York City	STI & TB clinics, Jails	OBS (baseline control)	Rapid HIV test (pamphlet & video pre-test counseling)	Standard HIV test (routine pre-test counseling)	NS	91292 (47771)	Test volume	Increased test volume (by 68%)	³¹

Study (Period, Location)	Setting	Study design	Intervention (POC test used)	Control	Eligibility	Sample size (Intervention arm)	Outcome	Result (Intervention compared to Control arms)	Ref
Kroc 2001 Chicago	ER	OBS (baseline control)	Rapid HIV test, routine offer (SUDS)	Standard HIV test, targeted offer	18-60 years HIV-	NS (1652)	Result turnaround time Test volume Result receipt	Increased turnaround time (average <2 hrs vs 14.5 days) Increased test volume (9.5 times greater). Increased receipt of results (99.3% vs 24%)	³⁹
Greenwald 2001-2003 Boston	Hospital (In-patients)	OBS (baseline control)	Rapid HIV test (OraQuick)	Non-rapid HIV test (OraSure) with results in 7 days	Inpatients	NS	Testing offer Result receipt	More patients offered testing (100% vs 50-66%). Note: also increased number of testing staff Increased receipt of results (100% vs 77%)	⁹
CDC 2000 Atlanta	Urgent care centre	OBS (baseline control)	Rapid HIV test (SUDS) or standard HIV testing, routine offer	Standard HIV test, targeted offer	18-65 years HIV- Last test > 6 mos	3887 (2787)	Test volume Detection of New HIV+ infections Result receipt (for HIV+) Entry to HIV care	Increased test volume (2.5 times, sig). Note: different recruitment strategies – not a valid comparison. More new HIV+ infections detected (1.6 times, sig), likely related to volume Increased receipt of HIV+ test result (2.0 times, sig) Increased entry into care (2 times, sig)	²⁸
Paneth-Pollak 2004-2005 New York City	STI clinics	OBS (concurrent control)	Rapid HIV test (OraQuick)	Standard HIV test	NS	23586 (21687)	Predictors of test choice Receipt of results	No difference in: HIV prevalence, demographics, concurrent STI diagnosis, most risk behaviours. Clients with rapid test more likely to report previous STI (18.3% vs 16.2%, p<0.02). Increased receipt (97.8% vs 63.3%, p<0.0001)	⁴³
Shima 2003 Japan	Public health center	OBS (baseline control)	Rapid HIV test (Unknown)	Standard HIV test	NS	583 (453)	Test volume STI testing	250% increase in test volume, with promotion of testing. Decreased proportion of clients tested for chlamydia (33% vs 76%) and syphilis (63% vs 77%). STI test volume increased due to overall increase in number of clients presenting.	⁴⁵
Landrum 2003 Texas	Occupational exposure	OBS (baseline control)	Rapid HIV test of source (OraQuick)	Standard HIV test	Source patients from occupational blood exposure	150 (79)	Result turnaround time Uptake of PEP Psychological distress	Faster turnaround time (mean 117 min vs 3254 min) Fewer mean doses PEP dispensed (5.5 vs 12.7; not sig), fewer mean doses PEP ingested (1.2 vs 3.8; sig). No large differences in self-reported anxiety, except fewer reported repeated thoughts of the exposure (26% vs 61%, p=0.049)	⁴⁰
Puro 1995-2002 Italy	Occupational exposure	OBS (baseline control)	Rapid HIV test of source (Capillus HIV-1/HIV-2)	Standard HIV test	Source patients from occupational blood exposure	1481 (769)	Source patients tested Uptake of PEP	Increase in proportion of source patients tested (88.6% vs 62.9%) Decrease in proportion of HCW started on PEP (2.3% vs 18.4%)	⁴⁴

Study (Period, Location)	Setting	Study design	Intervention (POC test used)	Control	Eligibility	Sample size (Intervention arm)	Outcome	Result (Intervention compared to Control arms)	Ref
Forsyth 2000 USA	Pregnant women	OBS (baseline control)	Rapid HIV test (SUDS)	Standard HIV test	NS	100 (36)	Result turnaround time Result receipt	Increased turnaround time (mean 8 hours vs 35.3 hours, p<0.0001) Greater availability of results available prior to rupture of membranes (36% vs 2%) and delivery (64% vs 8%)	²⁹
Garcia 2004-2005 Illinois	Pregnant women	OBS (baseline control)	Rapid HIV test	Standard HIV test	NS	NS (99888)	Documentation of HIV status	Prior to program implementation, 28% of women arrived in labour and delivery without a documented HIV status. After program implementation this proportion steadily decreased to 10% by Sept 2005 (sig).	³⁰
Hillis 2002, 2004 Russia	Pregnant women	OBS (baseline control)	Rapid HIV test (Determine HIV-1/HIV-2) (universal offer)	Standard HIV test (targeted offer)	HIV- or unknown No HIV test after 34 weeks GA	NS (704)	Receipt of ART	Greater receipt of ART (76% vs 41%).	³²
Bulterys 2001-2003 USA	Pregnant women	OBS (concurrent control)	Rapid HIV test (OraQuick)	Standard HIV test (simultaneous)	HIV status undocumented Active labour at ≥ 24 weeks GA Presentation ≥ 34 weeks GA	4849 (4849)	Result turnaround time	Decreased turnaround time for results (median 66 min [IQR 45-120 min] vs median 28 hours; p<0.001).	²⁷

Legend: ART (antiretroviral therapy), CBO (community-based organization), ER (Emergency room), GA (gestational age), NS (not specified), NEX (needle exchange), SUDS (Single Use Diagnostic System HIV-1 test)

REFERENCES

- (1) Centers for Disease Control and Prevention (CDC). Rapid HIV test distribution--United States, 2003-2005. *MMWR Morb Mortal Wkly Rep.* 2006;55:673-676.
- (2) Rekart ML, Krajden M, Cook D et al. Problems with the fast-check HIV rapid test kits. *CMAJ.* 2002;167:119.
- (3) Fonseca, K, Di Francesco L, Galli R, Hogg B, Schechter M, Kane S, Taylor D, Miller ML, Rekart M, Gill J, Read R, and George A. Results from a multi-centre Canadian clinical trial of a rapid HIV antibody test for use in point-of-care, clinical and laboratory settings. XV International AIDS Conference; 2004 Jul 11-16; Bangkok, Thailand.
- (4) BC Centre for Disease Control. Client counselling guidelines for use with point of care HIV test kits. 2007.
- (5) BC Centre for Disease Control. Guidelines for the use of point of care HIV test kits. 2007.
- (6) Health Canada. Proposed revisions to Point-of-Care HIV testing using rapid HIV test kits: guidance for health-care professionals (v 26-07, April 1 2000 CCDR, PPHB, Health Canada). 2007.
- (7) Greenwald JL, Burstein GR, Pincus J, Branson B. A rapid review of rapid HIV antibody tests. *Current Infectious Disease Reports.* 2007;8:125-131.
- (8) Sands A, Vercauteren G, Beelaert G, and Fransen K. HIV rapid testing for scaling up HIV prevention efforts. XVI International AIDS Conference; 2006 Aug 13-18; Toronto, Canada.
- (9) Greenwald JL, Hall J, Skolnik PR. Approaching the CDC's guidelines on the HIV testing of inpatients: physician-referral versus nonreferral-based testing. *AIDS Patient Care STDS.* 2006;20:311-317.
- (10) Centers for Disease Control and Prevention (CDC). Protocols for confirmation of reactive rapid HIV tests. *MMWR Morb Mortal Wkly Rep.* 2004;53:221-222.
- (11) Stekler J, Wood RW, Swenson PD, Golden M. Negative rapid HIV antibody testing during early HIV infection. *Ann Intern Med.* 2007;147:147-148.
- (12) Stekler, J, Swenson PD, Coombs RW, Dragavon J, Wood RW, and Golden MR. Anonymous testing & rapid testing in screening for acute HIV infection. 14th Conference on Retroviruses and Opportunistic Infections; 2007 Feb 25-28; Los Angeles, USA.
- (13) Klausner J, Philip S, Ahrens K, Nieri G, Robert R, Liska S, Louie B, Wong E, and Kent C. Combined rapid and HIV RNA testing improves case detection and potential prevention in a public STD clinic, San Francisco, 2004-2006. 14th Conference on Retroviruses and Opportunistic Infections; 2007 Feb 25-28; Los Angeles, USA.

- (14) Philip S, Ahrens K, Nieri G, Kohn R, Louie B, Liska S, Kent C, and Klausner J. State-of-the-art HIV testing - combined rapid and HIV RNA testing in a public STD clinic, San Francisco, 2004-2005. XVI International AIDS Conference; 2006 Aug 13-18; Toronto, Canada.
- (15) Delaney KP, Branson B, and Fridlund C. Ability of untrained users to perform rapid HIV antibody screening tests. American Public Health Association Annual Meeting; 2002 Nov 9-13; San Francisco, USA.
- (16) Kilby DL-P, Major CJ, Steben MH, Sutherland WD, and Bally GA. Point-of-care HIV testing using simple/rapid HIV test kits: guidance for health-care professionals. *Canada Communicable Disease Report*, 2000; 26-07:49-59.
- (17) Quality Assurance Guidelines for Testing Using Rapid HIV Antibody Tests Waived Under the Clinical Laboratory Improvement Amendments of 1988. Centers for Disease Control and Prevention, 2007.
- (18) Centers for Disease Control and Prevention (CDC). Guidelines for assuring the accuracy and reliability of HIV rapid testing: applying a quality system approach. World Health Organization, 2005.
- (19) AIDS Institute and Wadsworth Center and New York State Department of Health. Rapid HIV testing workbook and implementation guidelines for limited testing sites in New York State. 2005.
- (20) Branson B, Wesolowski L, Delaney K, Mavinkurve M, Dowling T, and MacKellar D. Investigation of reports of excessive false-positive oral fluid rapid HIV tests. 13th Conference on Retroviruses and Opportunistic Infections; Feb 5-8 2006; Denver, USA.
- (21) Delaney KP, Branson BM, Uniyal A et al. Performance of an oral fluid rapid HIV-1/2 test: experience from four CDC studies. *AIDS*. 2006;20:1655-1660.
- (22) Jafa K, Patel P, MacKellar DA et al. Investigation of false positive results with an oral fluid rapid HIV-1/2 antibody test. *PLoS ONE*. 2007;e185.
- (23) Javanbakht M, Uniyal A, Wigen C, Bolan B, Horton T, and Kerndt P. Evaluation of the accuracy of oral fluid rapid HIV tests after an unexpected number of false positive tests at a voluntary counseling and testing center. 14th Conference on Retroviruses and Opportunistic Infections; 2007 Feb 25-28; Los Angeles, USA.
- (24) Dowling, T and Facente S. False positive oral fluid rapid HIV tests in San Francisco prompt new testing protocol. XVI International AIDS Conference; 2006 Aug 13-18; Toronto, Canada.
- (25) Rekart ML, Quon JA, Rees T. Final evaluation results for the Fast-Check HIV rapid test kits. *CMAJ*. 2004;171:1324.
- (26) Antonio-Gaddy M, Richardson-Moore A, Burstein GR, Newman DR, Branson BM, Birkhead GS. Rapid HIV antibody testing in the New York State Anonymous HIV

- Counseling and Testing Program: experience from the field. *J Acquir Immune Defic Syndr*. 2006;43:446-450.
- (27) Bulterys MJD. Rapid HIV-1 testing during labor: a multicenter study.see comment. *JAMA*. 2004;292:219-223.
- (28) Centers for Disease Control and Prevention (CDC). Routinely recommended HIV testing at an urban urgent-care clinic--Atlanta, Georgia, 2000. *MMWR Morb Mortal Wkly Rep*. 2001;50:538-541.
- (29) Forsyth BW, Barringer SR, Walls TA et al. Rapid HIV testing of women in labor: too long a delay. *J Acquir Immune Defic Syndr*. 2004;35:151-154.
- (30) Garcia P, Bryant A, Olszewski Y, Statton A, and Cohen M. Reducing perinatal HIV transmission with rapid testing in labor and delivery: the Illinois experience. 13th Conference on Retroviruses and Opportunistic Infections; 2006 Feb 5-8; Denver, USA.
- (31) Henn M, Begier E, Sepkowitz KA, and Kellerman S. Less talk and more testing: how NYC has increased HIV testing. XVI International AIDS Conference; 2006 Aug 13-18; Toronto, Canada.
- (32) Hillis SD, Rakhmanova A, Vinogradova E et al. Rapid HIV testing, pregnancy, antiretroviral prophylaxis and infant abandonment in St Petersburg. *Int J STD AIDS*. 2007;18:120-122.
- (33) Hutchinson AB, Branson BM, Kim A, Farnham PG. A meta-analysis of the effectiveness of alternative HIV counseling and testing methods to increase knowledge of HIV status. *AIDS*. 2006;20:1597-1604.
- (34) Jackson P and Sloop N. The impact of rapid HIV testing on STD programs. XVI International AIDS Conference; 2006 Aug 13-18; Toronto, Canada.
- (35) Kassler WJ, Dillon BA, Haley C, Jones WK, Goldman A. On-site, rapid HIV testing with same-day results and counseling. *AIDS*. 1997;11:1045-1051.
- (36) Kelen GD, Shahan JB, Quinn TC. Emergency department-based HIV screening and counseling: experience with rapid and standard serologic testing. *Ann Emerg Med*. 1999;33:147-155.
- (37) Kendrick SR, Kroc KA, Withum D, Rydman RJ, Branson BM, Weinstein RA. Outcomes of offering rapid point-of-care HIV testing in a sexually transmitted disease clinic. *J Acquir Immune Defic Syndr*. 2005;38:142-146.
- (38) King AM, Osterwalder JJ, Vernazza PL. A randomised prospective study to evaluate a rapid HIV-antibody assay in the management of cases of percutaneous exposure amongst health care workers. *Swiss Med Wkly*. 2001;131:10-13.

- (39) Kroc K, Kendrick S, Withum D, Couture E, Miller S, Zagorski B, Branson B, and Weinstein R. Rapid HIV testing in an emergency department. 2002 National STD Prevention Conference; 2002 Mar 5; San Diego, USA.
- (40) Landrum ML, Wilson CH, Perri LP, Hannibal SL, O'Connell RJ. Usefulness of a rapid human immunodeficiency virus-1 antibody test for the management of occupational exposure to blood and body fluid. *Infect Control Hosp Epidemiol.* 2005;26:768-774.
- (41) Liang TS, Erbeding E, Jacob CA et al. Rapid HIV testing of clients of a mobile STD/HIV clinic. *AIDS Patient Care STDS.* 2005;19:253-257.
- (42) Metcalf CA, Douglas JM, Jr., Malotte CK et al. Relative efficacy of prevention counseling with rapid and standard HIV testing: a randomized, controlled trial (RESPECT-2). *Sex Transm Dis.* 2005;32:130-138.
- (43) Paneth-Pollak R, Mavinkurve M, Rubin S, Schillinger JA, and Blank S. Rapid versus conventional HIV testers; are there risk differences? 2006 National STD Prevention Conference; 2006 May 8-11; Jacksonville, USA.
- (44) Puro V, Francisci D, Sighinolfi L et al. Benefits of a rapid HIV test for evaluation of the source patient after occupational exposure of healthcare workers. *J Hosp Infect.* 2004;57:179-182.
- (45) Shima T, Isshiki M, Kondo M, Tsukada M, Shiomi S, Imai M. Introduction of rapid HIV tests to a public health center in Japan and evaluation of its effects. *Nippon Koshu Eisei Zasshi - Japanese Journal of Public Health.* 2006;53:167-177.
- (46) Shima T, Isshiki M, Tsukada M, Shiomi S, Yasunari R, Watanabe H, Ueyama H, Sudo K, Kondo M, Nakase K, and Imai M. Implementation and effectiveness of rapid HIV testing at publicly funded voluntary HIV counseling and testing (VCT) sites in Japan. XVI International AIDS Conference; 2006 Aug 13-18; Toronto, Canada.
- (47) Spielberg F, Branson BM, Goldbaum GM et al. Choosing HIV Counseling and Testing Strategies for Outreach Settings: A Randomized Trial. *J Acquir Immune Defic Syndr.* 2005;38:348-355.
- (48) Wurcel A, Zaman T, Zhen S, Stone D. Acceptance of HIV antibody testing among inpatients and outpatients at a public health hospital: a study of rapid versus standard testing. *AIDS Patient Care STDS.* 2005;19:499-505.
- (49) Guenter D, Greer J, Trow R, Browne G, Robinson G, and Roberts J. The effects of a rapid point-of-care HIV testing program: Hassle Free Clinic, Toronto. C03-3. Community-Linked Evaluation AIDS Resource. Working Paper Series. 2003.
- (50) Valdiserri RO, Moore M, Gerber AR, Campbell CH, Jr., Dillon BA, West GR. A study of clients returning for counseling after HIV testing: implications for improving rates of return. *Public Health Rep.* 1993;108:12-18.

- (51) Singh A, Hill S, Shafran S. Non-return rates for HIV testing: results of a 3-month retrospective review at an STD clinic in 1999. *Can Commun Dis Rep.* 2000;26:152-155.
- (52) Lazebnik R, Hermida T, Szubski R, Dieterich-Colon S, Grey SF. The proportion and characteristics of adolescents who return for anonymous HIV test results. *Sex Transm Dis.* 2001;28:401-404.
- (53) Kinsler JJ, Cunningham WE, Davis C, Wong MD. Time trends in failure to return for HIV test results. *Sex Transm Dis.* 2007;34:397-400.
- (54) Erbeding EJ, Chung S, Zenilman JM. Following-up for HIV test results: what limits return in an STD clinic population? *Int J STD AIDS.* 2004;15:29-32.
- (55) Marks G, Crepaz N, Senterfitt JW, Janssen RS. Meta-analysis of high-risk sexual behavior in persons aware and unaware they are infected with HIV in the United States: implications for HIV prevention programs. *Journal of Acquired Immune Deficiency Syndromes: JAIDS* 39(4):446-53. 2005.
- (56) Centers for Disease Control and Prevention (CDC). Update: HIV counseling and testing using rapid tests--United States, 1995. *MMWR Morb Mortal Wkly Rep.* 1998;47:211-215.
- (57) Winter AJ, Sulaiman Z, Hawkins D, British Association for Sexual Health and HIV Clinical Governance C. BASHH Clinical Governance Committee guidance on the appropriate use of HIV point of care tests. *Int J STD AIDS.* 2006;17:802-805.
- (58) Branson BM, Handsfield HH, Lampe MA et al. Revised recommendations for HIV testing of adults, adolescents, and pregnant women in health-care settings. *Morbidity & Mortality Weekly Report Recommendations & Reports* 55(RR-14):1-17; quiz CE1-4. 2006.
- (59) Centers for Disease Control and Prevention (CDC). Rapid point-of-care testing for HIV-1 during labor and delivery--Chicago, Illinois, 2002. *MMWR Morb Mortal Wkly Rep.* 2003;52:866-868.
- (60) Rajegowda BK, Das BB, Lala R, Rao S, Mc Neeley DF. Expedited human immunodeficiency virus testing of mothers and newborns with unknown HIV status at time of labor and delivery. *J Perinat Med.* 2000;28:458-463.
- (61) Kissin D, Akatova N, Rakhmanova A, Vinogradova E, Vornin E, Jamieson D, Glynn M, Robinson J, Miller W, and Hillis S. Rapid HIV testing and prevention of mother-to-child HIV transmission in high-risk maternity hospitals, St. Petersburg, Russia, 2004-2005. 13th Conference on Retroviruses and Opportunistic Infections; 2006 Feb 5-8; Denver, USA.
- (62) Van Dyke R, Scott G, Cohen M, Nesheim S, Abrams E, Danner S, and Bulterys M. The outcome of infants born to HIV-infected women identified by rapid HIV testing late in pregnancy or at delivery: the MIRIAD study, 2001 to 2004. 12th Conference on Retroviruses and Opportunistic Infections; 2005 Feb 22-25; Boston, USA.

- (63) Aaron E, Levine AB, Monahan K, Biondo CP. A rapid HIV testing program for labor and delivery in an inner-city teaching hospital. *AIDS Read*. 2006;16:22-24.
- (64) Csete J and Elliott R. Prevention and protection: enhancing both HIV testing and human rights in Canada. Canadian HIV/AIDS Legal Network . 2007.
- (65) Webber MP, Demas P, Enriquez E et al. Pilot study of expedited HIV-1 testing of women in labor at an inner-city hospital in New York City. *Am J Perinatol*. 2001;18:49-57.
- (66) Jamieson DJ, O'Sullivan MJ, Maupin R et al. The challenges of informed consent for rapid HIV testing in labor. *J Womens Health*. 2003;12:889-895.
- (67) Zlotnik-Shaul R. Rapid HIV testing in the delivery room; is a valid informed consent possible? 2003 Ontario HIV Treatment Network Research Conference; 2003 Nov 3-4; Toronto, Canada.
- (68) Stokes SH, McMaster P, Ismail KM. Acceptability of perinatal rapid point-of-care HIV testing in an area of low HIV prevalence in the UK. *Arch Dis Child*. 2007;92:505-508.
- (69) Monahan K, Aaron E, Williams T, and Wolf S. Rapid HIV testing as the standard of care in an urban OB/GYN clinic: lessons learned. 132nd Annual Meeting of the American Public Health Association; 2004 Nov 6-10; Washington, USA.
- (70) Fisher M, Benn P, Evans B, Pozniak A, Jones M, Maclean S, Davidson O, Summerside J, Hawkins D, and Clinical Effectiveness Group (British Association for Sexual Health and HIV). UK Guideline for the use of post-exposure prophylaxis for HIV following sexual exposure. *Int J STD AIDS*, 2006;17(2):81-92.
- (71) Panlilo, AL, Cardo DM, Grohskopf LA, Heneine W, Ross CS, and U.S.Public Health Service. Updated U.S. Public Health Service guidelines for the management of occupational exposures to HIV and recommendations for postexposure prophylaxis. *MMWR*, 2005;54[RR-9]:1-17.
- (72) Smith DK, Grohskopf LA, Black RJ, Auerbach JD, Veronese F, Struble KA, Cheever L, Johnson MJ, Paxton LA, Onorato IM, Greenberg AE, and U.S.Department of Health and Human Services. Antiretroviral postexposure prophylaxis after sexual, injection-drug use, or other nonoccupational exposure to HIV in the United States: recommendations from the U.S. Department of Health and Human Services. *MMWR*, 2005; 54[RR-2]:1-20.
- (73) BC Centre for Excellence in HIV/AIDS. Management of accidental exposure to HIV. 2006.
- (74) BC Centre for Disease Control. Blood and Body Fluid Exposure Management. 2005.
- (75) Newey C and Nwokolo N. Post-exposure prophylaxis following sexual intercourse. *Int J STD AIDS*, 2004; 15(12):844-846.

- (76) Weis KE, Duffus WA, Branson B, and Garnder L. Missed opportunities to identify HIV-infected individuals in South Carolina, Jan 2001-Dec 2005. 13th Conference on Retroviruses and Opportunistic Infections; 2006 Feb 5-8; Denver, USA.
- (77) Calderon Y, Leider J, Hailpern SM, Haughey M, Lombardi P, Huang D, Jordan N, Bijur PE, and Bauman LJ. Program B.R.I.E.F: high volume, rapid HIV testing in an urban ED. 4th IAS Conference on HIV pathogenesis, treatment and prevention; 2007 Jul 22-25; Sydney, Australia.
- (78) Centers for Disease Control and Prevention (CDC). Rapid HIV testing in emergency departments--three U.S. sites, January 2005-March 2006. *MMWR*, 2007; 56(24):597-601.
- (79) Jain, C, MacKay R, Wallach F, and Wyatt C. Acceptance of rapid HIV testing in medical inpatients. 14th Conference on Retroviruses and Opportunistic Infections; 2007 Feb 25-28; Los Angeles, USA.
- (80) Kendrick SR, Kroc KA, Couture E, Weinstein RA. Comparison of point-of-care rapid HIV testing in three clinical venues. *AIDS*. 2004;18:2208-2210.
- (81) Lyss SB, Branson BM, Kroc KA, Couture EF, Newman DR, Weinstein RA. Detecting unsuspected HIV infection with a rapid whole-blood HIV test in an urban emergency department. *J Acquir Immune Defic Syndr*. 2007;44:435-442.
- (82) Sharp V, Suh J, Arenas M, Tsveniashvili L, Mack T, and Maslow C. Physician-mediated HIV testing. 14th Conference on Retroviruses and Opportunistic Infections; 2007 Feb 25-28; Los Angeles, USA.
- (83) Silva A, Glick NR, Lyss SB et al. Implementing an HIV and Sexually Transmitted Disease Screening Program in an Emergency Department. *Ann Emerg Med*. 2007;49:564-572.
- (84) Smerd R, Marsh R, Holzman RS, Pearlman E, Hyde M, Rakower D, and Aberg JA. Implementing routine rapid HIV testing in a large public healthcare facility. 14th Conference on Retroviruses and Opportunistic Infections; 2007 Feb 25-28; Los Angeles, USA.
- (85) Hutchinson AB, Corbie-Smith G, Thomas SB, Mohanan S, del Rio C. Understanding the patient's perspective on rapid and routine HIV testing in an inner-city urgent care center. *AIDS Educ Prev*. 2004;16:101-114.
- (86) Lubelchek R, Kroc K, Hota B et al. The role of rapid vs conventional human immunodeficiency virus testing for inpatients: effects on quality of care. *Arch Intern Med*. 2005;165:1956-1960.
- (87) Mitterni L, Lin A, Stephen C, and Greer J. Effective delivery of HIV point of care testing in a major sexual health clinic. 15th Annual Canadian Conference on HIV/AIDS Research; 2006 May 25-28; Quebec City, Canada.

- (88) Rekart ML, Spencer DS, Knowles LJ, Barnett J, MacDougall R, Patrick DM, Cook D, MacLauchlan D, Cote YP, and Grevstad K. The impact on patients and care providers of a point of care rapid HIV test. BC Centre for Disease Control.
- (89) Smith LV, Rudy ET, Javanbakht M et al. Client satisfaction with rapid HIV testing: comparison between an urban sexually transmitted disease clinic and a community-based testing center. *AIDS Patient Care STDS*. 2006;20:693-700.
- (90) Padilla, S, Dilllon B, Iatesta M, Brooks L, Raveneau L, Malotte K, McKleroy V, and the RESPECT-2 Study Group. Counselor perceptions and impression of HIV counseling using a rapid HIV test: RESPECT-2. 2001 National HIV Prevention Conference; 2001 Aug 12-15; Atlanta, USA.
- (91) Hammett TM, Harmon MP, Rhodes W. The Burden of Infectious Disease Among Inmates of and Releasees From US Correctional Facilities, 1997. *American Journal of Public Health* 2002;92:1789-1794.
- (92) Rothon DA, Mathias RG, Schechter MT. Prevalence of HIV infection in provincial prisons in British Columbia.[see comment]. *CMAJ*, 1994;151(6):781-7.
- (93) MacGowan R, Margolis A, Richardson-Moore A et al. Voluntary rapid Human Immunodeficiency virus (HIV) testing in jails. *Sex Transm Dis* 2007; 34(11): published online ahead of print.
- (94) Beckwith CG, Atunah-Jay S, Cohen J et al. Feasibility and acceptability of rapid HIV testing in jail. *AIDS Patient Care STDS*. 2007;21:41-47.
- (95) Spaulding AC, Jacob Arriola KR, Hammett T, Kennedy S, Tinsley M. Rapid HIV Testing In Rapidly Released Detainees: Next Steps. *Sex Transm Dis* 2007; 34(11): published online ahead of print.
- (96) Centers for Disease Control and Prevention (CDC). Rapid HIV testing among racial/ethnic minority men at gay pride events--nine U.S. cities, 2004-2006. *MMWR*, 2007; 56(24):602-4.
- (97) Keenan PA, Keenan JM. Rapid hiv testing in urban outreach: a strategy for improving posttest counseling rates. *AIDS Educ Prev*. 2001;13:541-550.
- (98) Brunings P, Spencer D, Ogilvie G, Knowles L, Remple VP, Johnston C, Thindal S, Bungay V, and Patrick DM. Feasibility of and responses to point of care (POC) STI testing and rapid HIV screening by outreach workers for indoor commercial sex workers (CSW) in Vancouver, Canada. 17th Meeting of the International Society for Sexually Transmitted Diseases; 2007 Jul 29 - Aug 1; Seattle, USA.
- (99) Tyndall MW, Shannon K, Bright V, Prasad D, Lai C, Zhang R, and Kerr T. Offering rapid HIV testing to marginalized sex workers in Vancouver. 16th Annual Canadian Conference on HIV/AIDS Research; 2007 Apr 26-29; Toronto, Canada.

- (100) Bucher JB, Thomas KM, Guzman D, Riley E, Dela Cruz N, Bangsberg DR. Community-based rapid HIV testing in homeless and marginally housed adults in San Francisco. *HIV Medicine*. 2007;8:28-31.
- (101) Foster J, Vibert Y, Conway D, Chen J, Laguerre R, Van Kust B, Robertson C, Parrino T, and Boswell E. Rapid testing for children and youth: results from a pilot project in Philadelphia. 12th Conference on Retroviruses and Opportunistic Infections; 2005 Feb 22-25; Boston, USA.
- (102) Torres A, Loaiza L, Graves E, Lindahl MG, Lopez-Cevallos D, and Rink E. Improving access to HIV testing and counseling and prevention education: A rural Hispanic/Latino Promotores Program. 134th American Public Health Association Annual Meeting; 2006 Nov 4-8; Boston, USA.
- (103) Thompson M, Williams S, White E, Williams K, and Daniels E. Rapid HIV testing in the field reaches young and lower-income African-American women otherwise unlikely to be tested: the AIDS Research Consortium of Atlanta (ARCA) Metro Atlanta Women of Color Initiative (MAWOCI). XVI International AIDS Conference; 2006 Aug 13-18; Toronto, Canada.
- (104) Paul S, Cadoff E, Wolski M, Martin RM, Maung A, Berezny L, Nichol L, Harvey-Talbot M, Rhonda W, Drumm S, Salaru G, and Martin E. Rapid HIV testing on mobile vans. XVI International AIDS Conference; 2006 Aug 13-18; Toronto, Canada.
- (105) Smith MJ, Wolff K, and Kostman J. Implementation of rapid HIV-1 antibody testing in an inpatient drug and alcohol recovery setting. XVI International AIDS Conference; 2006 Aug 13-18; Toronto, Canada.
- (106) Pugatch DL, Levesque BG, Lally MA et al. HIV testing among young adults and older adolescents in the setting of acute substance abuse treatment. *J Acquir Immune Defic Syndr*. 2001;27:135-142.
- (107) Prost A, Chopin M, Mcowan A et al. "There is such a thing as asking for trouble": taking rapid HIV testing to gay venues is fraught with challenges. *Sex Trans Inf*, 2007; 83(3):185-8.
- (108) Elliott R and Jurgens R. Rapid HIV screening at the point of care: legal and ethical questions. Canadian HIV/AIDS Legal Network. 2000.